

SEEMORE™ USB PROBE SPECIFICATIONS

USB CABLES

Probes come with a 2 meter USB cable. 5 meter cable optional.

MINIMUM SYSTEM REQUIREMENTS

- Computer Operating System: Windows XP, Vista, or Mac (running Windows)
- Processor: 1 GHz Minimum
- Memory: 512 K (recommended 1 GB or more)
- USB 2.0 port
- Minimum Display: 1024X 600 with 32 bit color

SEEMORE™ USER INTERFACE

Intuitive and easy to use, the SeeMore™ system requires minimum input from the user to produce a desirable image. Software functions include:

- AutoScan – presets, customizable
- Manual Scan (DGC, Power Settings)
- Zoom
- User selectable depth, frequency
- Frame rate: 10-20 FPS, depth dependent
- Measurements, arrows, annotation: distance, area
- Calculations (optional): OB, Bladder Volume, Prostate volume
- Image store and review (backscatter images and/or JPG)
- Study Report (optional)
- Interface to EMR/HIS/PACS (optional)
- Freeze image with button on probe, space bar, or on-screen button
- Cine Loop
- Image averaging, interpolation
- 256 shades of gray
- Biopsy/needle path guideline (optional)